



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
 DIVISION OF PUBLIC HEALTH, STATE LABORATORY OF PUBLIC HEALTH

ROY COOPER  
 GOVERNOR

MANDY COHEN, MD, MPH  
 SECRETARY

DANIEL STALEY  
 DIRECTOR

SCOTT J. ZIMMERMAN, DRPH, MPH, HCLD (ABB)  
 LABORATORY DIRECTOR

**Memo**

**To:** Evelyn Foust  
 Communicable Disease Branch Head

**From:** Heather Colvin, B.S  
 Laboratory Improvement Consultant

**CC:** Pete Moore  
 Ron Higginbotham  
 NCSLPH Lab Improvement Staff  
 Local Health Department Lab Managers

**Date:** June 5th, 2017

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

**GC-Lect:** Lot Number: **7130741** Expiration Date: **08/18/2017**  
 Along with **Bio-Bag™ Type C:** Lot Number: **6188936** Expiration Date: **04/30/2018**  
 and with **Bio-Bag™ Type C:** Lot Number: **6257773** Expiration Date: **07/31/2018**

*Summary of NCSLPH QC results for this Lot #*

N. gonorrhoeae ATCC# 43069 - Growth	S. epidermidis ATCC# 12228 – No Growth
N. gonorrhoeae ATCC# 43070 - Growth	P. mirabilis ATCC# 43071 – Growth*
Sterility Check - Acceptable	Visual Check - Acceptable

\*While performing quality control testing on the GC-Lect agar plates, there was a breakthrough growth of *Proteus mirabilis* (gram negative rod) at **24 hours**. Please pay attention to possible growth of other organisms, especially at the **24 hour** mark. Follow your laboratory’s procedure for the presumptive identification of *Neisseria gonorrhoeae*.

**Lot number failed:** None

BD and McKesson® contacted: June 2nd, 2017

\*If your contact information changes, please provide that information to Heather Colvin in Laboratory Improvement at (919) 733-7186 or via e-mail at [heather.colvin@dhhs.nc.gov](mailto:heather.colvin@dhhs.nc.gov). Thank you.