



**ROY COOPER** • Governor  
**KODY H. KINSLEY** • Secretary  
**MARK BENTON** • Deputy Secretary for Health  
**SUSAN KANSAGRA MD, MBA** • Assistant Secretary for Public Health  
 Division of Public Health  
**SCOTT M. SHONE, PhD, HCLD (ABB)** • Laboratory Director, State  
 Laboratory of Public Health

# Memo

**To:** Evelyn Foust  
 Communicable Disease Branch Head

**From:** Lettia Spruill  
 Medical Laboratory Technologist I

**CC:** Pete Moore  
 Ron Higginbotham  
 NCSLPH Lab Improvement Staff  
 Local Health Department Lab Managers

**Date:** February 13, 2023

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

**Modified Thayer Martin II:** Lot Number: **150430** Expiration Date: **3/31/2023**  
 Along with **Bio-Bag™ Type C:** Lot Number: **B01E200M** Expiration Date: **2/28/2024** and **Bio-Bag™ Type C:** Lot Number: **B01E182M** Expiration Date: **2/28/2024**

**Summary of NCSLPH QC results for this Lot #:**

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
	<i>P. mirabilis</i> ATCC# 43071 – 1 colony at 24 hrs
Sterility Check - Acceptable	Visual Check - Acceptable

**Lot number failed:** None  
 Hardy Diagnostics Modified Thayer Martin

\*\*If your contact information changes, please provide that information to Linda Cleve in Laboratory Improvement at (252) 414-3078 or via e-mail at Linda.cleve@dhhs.nc.gov. Thank you.