



NC DEPARTMENT OF
**HEALTH AND
HUMAN SERVICES**
Division of Public Health

ROY COOPER • Governor

MANDY COHEN, MD, MPH • Secretary

MARK T. BENTON • Assistant Secretary for Public Health

Division of Public Health

SCOTT M. SHONE, PhD, HCLD (ABB) • Laboratory Director, State
Laboratory of Public Health

Memo

To: Evelyn Foust
Communicable Disease Branch Head

From: Angela Bradley, MLT (ASCP)
Laboratory Improvement Medical Laboratory Technologist

CC: Pete Moore
Ron Higginbotham
NCSLPH Lab Improvement Staff
Local Health Department Lab Managers

Date: May 13th, 2021

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Modified Thayer Martin II: Lot Number: 483532 Expiration Date: 07/01/2021
Along with **Bio-Bag™ Type C:** Lot Number: B02C099M Expiration Date: 10/30/2021
and with **Bio-Bag™ Type C:** Lot Number: B03C342M Expiration Date: 9/30/2022

Summary of NCSLPH QC results for this Lot #:

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
	<i>P. mirabilis</i> ATCC# 43071 –No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

Lot number failed: None
Hardy Diagnostics Modified Thayer Martin

**If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at angela.bradley@dhhs.nc.gov. Thank you.