



ROY COOPER • Governor
KODY H. KINSLEY • Secretary
MARK T. BENTON • Deputy Secretary for Health
SUSAN KANSAGRA MD, MBA Assistant • Secretary for Public Health
Division of Public Health
SCOTT M. SHONE, PhD, HCLD (ABB) • Laboratory Director, State
Laboratory of Public Health

Memo

To: Evelyn Foust
Communicable Disease Branch Head

From: Aja Stanback, BS
Quality Improvement Consultant

CC: Pete Moore
Ron Higginbotham
NCSLPH Lab Improvement Staff
Local Health Department Lab Managers

Date: July 28, 2022

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Modified Thayer Martin II: Lot Number: **509635** Expiration Date: **9/15/2022**
Along with **Bio-Bag™ Type C:** Lot Number: **B03C342M** Expiration Date: **9/30/2022**

Summary of NCSLPH QC results for this Lot #:

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
	<i>P. mirabilis</i> ATCC# 43071 –No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

Lot number failed: None
Hardy Diagnostics Modified Thayer Martin

**If your contact information changes, please provide that information to Linda Cleve in Laboratory Improvement at (252) 414-3078 or via e-mail at Linda.cleve@dhhs.nc.gov. Thank you.