



ROY COOPER • Governor
KODY H. KINSLEY • Secretary
MARK BENTON • Deputy Secretary for Health
SUSAN KANSAGRA MD, MBA • Assistant Secretary for Public Health
 Division of Public Health
SCOTT M. SHONE, PhD, HCLD (ABB) • Laboratory Director, State
 Laboratory of Public Health

Memo

To: Evelyn Foust
 Communicable Disease Branch Head

From: Lettia Spruill
 Medical Laboratory Technologist I

CC: Pete Moore
 Ron Higginbotham
 NCSLPH Lab Improvement Staff
 Local Health Department Lab Managers

Date: December 13, 2023

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Modified Thayer Martin II: Lot Number: **615707** Expiration Date: **01/29/2024**
 Along with **Bio-Bag™ Type C:** Lot Number: **B01E182M** Expiration Date: **02/28/2024**

Summary of NCSLPH QC results for this Lot #:

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
	<i>P. mirabilis</i> ATCC# 43071 – No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

Lot number failed: None
 Hardy Diagnostics Modified Thayer Martin

**If your contact information changes, please provide that information to Tracey Shives in Laboratory Improvement at (336)-306-4302 or via email at tracey.shives@dhhs.nc.gov. Thank you.