



NC DEPARTMENT OF  
**HEALTH AND  
HUMAN SERVICES**  
Division of Public Health

ROY COOPER • Governor  
MANDY COHEN, MD, MPH • Secretary  
DANNY STALEY • Director, Division of Public Health  
SCOTT J. ZIMMERMAN, DrPH, MPH, HCLD (ABB) • Laboratory  
Director, State Laboratory of Public Health

# Memo

**To:** Evelyn Foust  
Communicable Disease Branch Head

**From:** Angela Bradley, MLT (ASCP)  
Medical Laboratory Technologist

**CC:** Pete Moore  
Ron Higginbotham  
NCSLPH Lab Improvement Staff  
Local Health Department Lab Managers

**Date:** September 17<sup>th</sup>, 2018

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

**Modified Thayer Martin II:** Lot Number: **8228592** Expiration Date: **12/14/2018**  
Along with **Bio-Bag™ Type C:** Lot Number: **8129674** Expiration Date: **01/31/2020**  
and with **Bio-Bag™ Type C:** Lot Number: **7296619** Expiration Date: **06/30/2019**

**Summary of NCSLPH QC results for this Lot #:**

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
<i>N. gonorrhoeae</i> ATCC# 43070 - Growth	<i>P. mirabilis</i> ATCC# 43071 – No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

**Lot number failed:** None  
BD and Fisher Scientific contacted: September 13<sup>th</sup>, 2018.

\*If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at [angela.bradley@dhhs.nc.gov](mailto:angela.bradley@dhhs.nc.gov). Thank you.

NC DEPARTMENT OF HEALTH AND HUMAN SERVICES • DIVISION OF PUBLIC HEALTH • STATE LABORATORY OF PUBLIC HEALTH

LOCATION: 4312 District Drive, Raleigh, NC 27607

MAILING ADDRESS: 1918 Mail Service Center, Raleigh, NC 27699-1918

[www.ncdhhs.gov](http://www.ncdhhs.gov) • <http://publichealth.nc.gov> • [slph.ncpublichealth.com](http://slph.ncpublichealth.com) • TEL: 919-733-7834 • FAX: 919-733-8695

AN EQUAL OPPORTUNITY / AFFIRMATIVE ACTION EMPLOYER