



NC DEPARTMENT OF
**HEALTH AND
HUMAN SERVICES**
Division of Public Health

ROY COOPER • Governor

MANDY COHEN, MD, MPH • Secretary

BETH LOVETTE, MPH, BSN, RN • Acting Director,
Division of Public Health

SCOTT J. ZIMMERMAN, DrPH, MPH, HCLD (ABB) • Laboratory
Director, State Laboratory of Public Health

Memo

To: Evelyn Foust
Communicable Disease Branch Head

From: Angela Bradley, MLT (ASCP)
Medical Laboratory Technologist

CC: Pete Moore
Ron Higginbotham
NCSLPH Lab Improvement Staff
Local Health Department Lab Managers

Date: June 17th, 2019

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Modified Thayer Martin II: Lot Number: **9143726** Expiration Date: **9/20/2019**
Along with **Bio-Bag™ Type C:** Lot Number: **8129674** Expiration Date: **01/31/2020**
and with **Bio-Bag™ Type C:** Lot Number: **7296619** Expiration Date: **06/30/2019**

Summary of NCSLPH QC results for this Lot #:

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
<i>N. gonorrhoeae</i> ATCC# 43070 - Growth	<i>P. mirabilis</i> ATCC# 43071 – No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

Lot number failed: None
BD and Fisher Scientific contacted: June 14th, 2019.

*If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at angela.bradley@dhhs.nc.gov. Thank you.